

ECISS
 EUROPEAN COMMITTEE FOR IRON AND STEEL STANDARDISATION
 COMITÉ EUROPÉEN DE NORMALISATION DU FER ET DE L'ACIER
 EUROPÄISCHES KOMITEE FÜR EISEN- UND STAHLNORMUNG

EUROPEAN CERTIFIED REFERENCE MATERIAL (EURONORM – CRM)
 CERTIFICATE OF CHEMICAL ANALYSIS

EURONORM – CRM No. 056-2 0.8% Carbon Steel

LABORATORY MEANS (4 Values)
 mass content in %

Line No	C	Si	Mn	P	S	Cr	Ni	Al (Acid sol)	Cu	Al (Tot)
1	0.8087	—	—	—	0.0083	0.0139	—	0.00010	0.0120	<0.0001
2	0.8126	0.1901	0.5028	0.0090	0.0084	0.0140	0.0202	0.00010	0.0121	0.0002
3	0.8130	0.1948	0.5040	0.0094	0.0084	0.0141	0.0206	0.00015	0.0124	0.0002
4	0.8133	0.1963	0.5043	0.0099	0.0085	0.0142	0.0208	0.00019	0.0125	0.0003
5	0.8146	0.1980	0.5046	0.0100	0.0089	0.0142	0.0211	0.00020	0.0128	0.0003
6	0.8148	0.1982	0.5061	0.0100	0.0090	0.0143	0.0213	0.00024	0.0129	0.0005
7	0.8157	0.1983	0.5066	0.0100	0.0091	—	0.0214	0.00026	0.0129	0.0006
8	0.8175	0.1996	0.5067	0.0103	0.0092	0.0145	0.0216	0.00030	0.0130	0.0008
9	0.8192	0.2010	0.5072	0.0103	0.0094	0.0145	0.0217	0.00033	0.0130	0.0008
10	0.8194	0.2015	—	0.0103	0.0095	0.0145	0.0219	0.00035	0.0130	0.0009
11	0.8196	0.2017	0.5077	0.0104	0.0096	0.0145	0.0220	0.00040	0.0131	
12	0.8205	0.2026	0.5079	0.0105	0.0096	0.0146	0.0221	—	0.0131	
13	0.8211	0.2027	0.5082	0.0105	0.0098	0.0146	0.0221		0.0132	
14	0.8229	0.2041	0.5089	0.0107	0.0099	0.0149	0.0222		0.0132	
15	0.8237	0.2043	0.5097	0.0107	0.0099	0.0150	0.0222		0.0133	
16	0.8259	0.2045	0.5111	0.0109	0.0100	0.0152	0.0223		0.0133	
17	0.8260	0.2048	0.5133	0.0114	0.0103	0.0152	0.0224		0.0133	
18		0.2079	—	—		0.0153	0.0225		0.0133	
19				0.0115		0.0156	0.0235		—	
M _M	0.8181	0.2006	0.5073	0.0103	0.0093	0.0146	0.0218	0.00024	0.0129	
S _M	0.0050	0.0044	0.0029	0.0007	0.0007	0.0005	0.0008	0.00011	0.0005	
S _W	0.0025	0.0024	0.0021	0.0004	0.0002	0.0003	0.0003	0.00004	0.0003	

M_M: Mean of the intralaboratory means, S_M: standard deviation of the intralaboratory means, S_W: Intralaboratory standard deviation

The laboratory mean values have been examined statistically to eliminate outlying values. Where a "—" appears in the table it indicates that an outlying value has been eliminated by either the Cochran or Grubbs Test.

Additional Information from one laboratory: Mo: 0.0015%, Co: 0.0035%, N: 0.0045%

CERTIFIED VALUES
 Mass content in %

	C	Si	Mn	P	S	Cr	Ni	Al (Acid sol)	Cu
M _M	0.8181	0.2006	0.5073	0.0103	0.0093	0.0146	0.0218	0.00024	0.0129
C(95%)	0.0026	0.0023	0.0016	0.0004	0.0004	0.0003	0.0004	0.00007	0.0003

The half-width confidence interval C(95%) = $\frac{t \times S_M}{\sqrt{n}}$ where "t" is the appropriate Student's t value and "n" is the number of acceptable mean values

For further information regarding the confidence interval for the certified value see ISO Guide 35:2006 sections 6.1 and 10.5.2

This reference material was prepared in accordance with the recommendations set
 out in ISO Guides 30 – 35 and issued by:

BUREAU OF ANALYSED SAMPLES LIMITED
 Newham Hall, Middlesbrough, England TS8 9EA

On behalf of:- The Iron and Steel Nomenclature Co-ordinating Committee (COCOR) of the ECISS, after approval by all the participating laboratories and all the producing organizations. (France – IRSID/CTIF, Germany – Iron and Steel CRM Working Group: VDEh, BAM & MPI für Eisenforschung, Nordic Countries – Nordic CRM Working Group, UK – BAS Ltd.)



4004

METHODS USED
EURONORM – CRM No. 056-2

Element	Line Number	Methods
C	1-3-4-5-6-8-9-10-12-13-14-15-16-17	Combustion, infrared absorption
	2	Combustion, gravimetry
	7	Combustion, non-aqueous titration after absorption in organic solvent
	11	Combustion, thermal conductivity
Si	2-3-7-9-13-18	Inductively coupled plasma - optical emission spectrometry
	4-5-6-8-10-11-12-14-15	Gravimetry, dehydration with perchloric acid
	16-17	Spectrophotometry, molybdenum blue without extraction
Mn	2-7-8	Flame atomic absorption spectrometry
	3-6-12-13-16	Spectrophotometry, periodate oxidation
	4-5-9-11-14-15-17	Inductively coupled plasma - optical emission spectrometry
P	2-7-10-16	Spectrophotometry, molybdenum blue, without extraction
	3-4-6-8-13-17-19	Inductively coupled plasma - optical emission spectrometry
	5-9-11-12-14-15	Spectrophotometry, phosphovanadomolybdate, extraction
S	1-2-3-4-5-6-7-8-9-10-11-12-13-15-17	Combustion, infrared absorption
	14	Gravimetry as barium sulphate without separation
	16	Combustion, acidimetric titration, absorption in H ₂ O ₂ or AgNO ₃
Cr	1-2-3-15-16	Flame atomic absorption spectrometry
	4-5-6-8-9-10-11-12-13-14-17-18-19	Inductively coupled plasma - optical emission spectrometry
Ni	2-3-7-8-9-10-11-13-15-16-18-19	Inductively coupled plasma - optical emission spectrometry
	4-5-6-12-14	Flame atomic absorption spectrometry
	17	Spectrophotometry, dimethylglyoxime, without extraction
Al (Acid sol)	1-6	Inductively coupled plasma - mass spectrometry
	2-3-4-5-7-8-9-10-11	Inductively coupled plasma - optical emission spectrometry
Cu	1-5-6-7-9-10-12-17-18	Inductively coupled plasma - optical emission spectrometry
	2-3-4-8-11-13-14-16	Flame atomic absorption spectrometry
	15	Spectrophotometry, diethyldithiocarbamate, with extraction
Al (Tot)	1-3-4-5-6-8-9-10	<i>Inductively coupled plasma - optical emission spectrometry</i>
	2	<i>Glow discharge mass spectrometry</i>
	7	<i>Inductively coupled plasma - mass spectrometry</i>

PARTICIPATING LABORATORIES

AG der Dillinger Hüttenwerke, Dillingen, Germany
 Allvac Ltd., Sheffield, UK.
 ArcelorMittal, Dunkerque, France
 ArcelorMittal, Florange, France
 BAM Bundesanstalt für Materialforschung und -prüfung, Berlin, Germany
 Böhler Edelstahl GmbH, Kapfenberg, Austria
 Carsid S.A., Marcinelle, Belgium
 Centro Nacional de Investigaciones Metalúrgicas (CENIM), Madrid, Spain
 Corus Stahl, IJmuiden, Holland
 Corus Strip Products, Llanwern, UK
 Corus Testing Solutions, Scunthorpe, UK

Deutsche Edelstahlwerke GmbH, Siegen, Germany
 Deutsche Edelstahlwerke GmbH, Witten, Germany
 Gustavsson Consulting, Stockholm, Sweden
 Kanthal AB, Hallstahammar, Sweden
 Luxcontrol SA, Esch-sur-Alzette, Luxembourg
 Ovako Steel AB, Hofors, Sweden
 Pattinson and Stead (2005) Ltd., Middlesbrough, UK
 Ridsdale & Co Ltd., Middlesbrough, UK
 Swerea KIMAB, Stockholm, Sweden
 Uppsala University, Uppsala, Sweden
 voestalpine Stahl GmbH, Linz, Austria

DESCRIPTION OF THE SAMPLE

The sample consists of chips passing a nominal 1700µm aperture sieve from which the fines passing a nominal 250µm sieve has been removed. It is supplied in bottles containing 100g, ref ECRM 056-2(C). It is also supplied in the form of 44mm dia. discs, ref ECRM 056-2(D)

INTENDED USE & STABILITY

The chip sample, ECRM 056-2(C), is intended for the verification of analytical methods, such as those used by the participating laboratories, for the calibration of analytical instruments in cases where the calibration with primary substances (pure metals or stoichiometric compounds) is not possible and for establishing values for secondary reference materials.

It will remain stable provided that the bottle remains sealed and is stored in a cool, dry atmosphere. When the bottle has been opened the lid should be secured immediately after use. If the contents should become discoloured (e.g. oxidised) by atmospheric contamination they should be discarded.

The disc sample, ECRM 056-2(D), is intended for establishing and checking the calibration of Optical Emission and X-Ray Spectrometers for the analysis of similar materials. The "as received" working surface of the sample should be finished before use to remove any protective coating. The disc will remain stable provided that it is not subject to excessive heat (e.g. during preparation of the working surface).

TRACEABILITY

The traceability of ECRM 056-2 has been established in accordance with principles of ISO Guides 30 – 35 and the International vocabulary of basic and general terms in metrology.

The characterisation of this material has been achieved by inter-laboratory study, each laboratory using the method of their choice, details of which are given above. These methods are either stoichiometric analytical techniques or methods which are calibrated against pure metals or stoichiometric compounds. Most methods used were either international or national standard methods or methods which are technically equivalent.

FURTHER INFORMATION

For information regarding the preparation, certification and supply of these European Certified Reference Materials (EURONORM-CRMs) and the use of the statistical information given on this certificate, please refer either to the producer of this Certified Reference Material or to CEN Report CR 10317 and ECISS Information Circular No. 5, both of which are available from the national standards body in your country or from CEN in Brussels. (In the UK this is the BSI, 389 Chiswick High Road, London W4 4AL).

Further information and advice on this or other Certified Reference Materials or Reference Materials produced by Bureau of Analysed Samples Ltd. may be obtained from the address below.

Pour disposer d'informations sur la fabrication, la certification et la distribution des Matériaux de Référence Certifiés Européens (EURONORM-MRC) ainsi que sur l'utilisation des informations statistiques données sur ce certificat, se reporter soit au producteur de ce Matériau de Référence Certifié, soit au Rapport CEN CR 10317 et à la circulaire d'information No. 5 (ECISS). On peut se procurer ces deux documents auprès des organismes nationaux de normalisation ou auprès du CEN, Bruxelles. (Pour la France: AFNOR, 11 Avenue Francis de Pressensé, 93571 – St Denis la Plaine Cedex).

D'autres informations et avis au sujet de ce Matériau de Référence Certifié, ou de tout autre Matériau de Référence Certifié ou Matériau de Référence produits par le Bureau of Analysed Samples Ltd, peuvent être demandés en contactant l'adresse figurant dans le bas de ce Certificat.

Angaben über Herstellung, Zertifizierung und Bezugsmöglichkeiten dieser Europäischen Zertifizierten Referenzmaterialien (EURONORM-ZRM) sowie über die Anwendungen der in diesem Zertifikat enthaltenen statistischen Daten sind erhältlich beim Hersteller dieses zertifizierten Referenzmaterials, dessen Adresse auf diesem Zertifikat angegeben ist oder sie finden sich im CEN-Report CR 10317 und in der Mitteilung Nr. 5 (ECISS), beide zu beziehen durch die nationalen Normenorganisationen oder direkt von CEN, Brüssel. (In Deutschland bei der Vertriebsstelle des DIN: Beuth-Verlag GmbH, Burggrafenstrasse 4-10, 10787 Berlin).

Weitere Informationen und Hinweise zu diesem oder anderen durch Bureau of Analysed Samples Ltd hergestellten zertifizierten Referenzmaterialien oder Referenzmaterialien können unter der unten angegebenen Adresse erhalten werden.

För information angående tillverkning, certifiering och distribuering av dessa europeiska certifierade referensmaterial (EURONORM CRM) och för användning av statistisk information, som angivits i detta certifikat, refereras antingen till producenten av detta certifierade referensmaterial med angiven adress på certifikatet eller till CEN-rapport CR 10317 och Informationscirkulär Nr 5 (ECISS), som kan erhållas från den nationella standardiseringsorganisationen eller från CEN, Bryssel. (Sverige: SIS, S:t Paulsgatan 6, SE-118 80 Stockholm, Finland: SFS, PL. 116, FIN-002 41, Helsingfors, Danmark: DS, Kollegievej 6, DK-Charlottenlund 2920, Norge: NSF, Drammensveien, 145 A, Postboks 353 Skøyen, NO-0213 Oslo, Island: STRI, Holtagardar, IS-104 Reykjavik).

Ytterligare information och rådfrågan om detta eller andra Certifierade Referensmaterial/Referensmaterial, producerade av Bureau of Analysed Samples Ltd kan erhållas från angiven adress enligt nedan.

NEWHAM HALL, NEWBY,
MIDDLESBROUGH, ENGLAND, TS8 9EA
Email: enquiries@basrid.co.uk
Website: www.basrid.co.uk

For BUREAU OF ANALYSED SAMPLES LTD.

R P MEERES,
MANAGING DIRECTOR