

date

The Reference Material (BAM-E001) delivered by BAM, Division 7.5 "Technical Properties of Polymeric Materials", is prepared for test purposes and tested in accordance with DIN EN ISO/IEC 17025:2005.

The Rubber Test Sheets ID No.

| | | | | | | | |
|--------|--------|--------|--------|--------|--------|--------|--------|
| 4x81-1 | 48x1-2 | 48x1-3 | 4881-4 | 48x1-5 | 4x81-6 | 48x1-7 | 4881x8 |
| 48x2-1 | 48x2-2 | 4x82-3 | 48x2-4 | x882-5 | 4x82-6 | 4x82-7 | 488x-8 |
| 48x3-1 | 4x83-2 | x883-3 | 4x83-4 | 4x83-5 | 48x3-6 | 48x3-7 | 48x3-8 |
| 4x84-1 | 48x4-2 | 48x4-3 | 488x-4 | x884-5 | 4x84-6 | 4x84-7 | x884-8 |
| x885-1 | 4x85-2 | 48x5-3 | 488x-4 | 488x-5 | 48x5-6 | 4x85-7 | x885-8 |
| 4x86-1 | 488x-2 | 488x-3 | 48x6-4 | 4x86-5 | 4x86-6 | 488x-7 | x886-8 |
| x887-1 | 4x87-2 | 48x7-3 | 4x87-4 | x887-5 | 4x87-6 | 48x7-7 | 4x87-8 |
| 4x88-1 | 48x8-2 | 48x8-3 | 48x8-4 | 4x88-5 | 4x88-6 | 48x8-7 | x888-8 |
| 48x9-1 | x889-2 | x889-3 | 4x89-4 | 48x9-5 | 488x-6 | 48x9-7 | 4x89-8 |

are in conformance with ISO 4649, Annex B, clause B.2.

The test sheets are stamped with the following marks:

| | |
|-----------------------------------|-------------------------|
| BAM | (producer) |
| ISO 4649 | (standard) |
| Standard reference compound No. 1 | (type of elastomer) |
| BAM-E001 | (product number) |
| XXXX | (identification number) |
| 15 | (year of production) |

This rubber test sheet corresponding to ISO 4649, Annex B, clause B.2. is identical with the term "Vergleichselastomer" mentioned in DIN 53516.

The sheets are produced: (month, year).

by order

Bundesanstalt für Materialforschung und -prüfung (BAM)

Dietmar Schulze; Quality Manager
Division 7.5
Unter den Eichen 87
12205 Berlin
Federal Republic of Germany

Tel.: +49 30 8104-3340

Fax: +49 30 8104-1707

Email: crm-elastomer@bam.de

Webshop: <https://www.webshop.bam.de/>