

date

The Reference Material (BAM-E005) delivered by BAM, Division 7.5 "Technical Properties of Polymeric Materials", is prepared for test purposes and tested in accordance with DIN EN ISO/IEC 17025:2005.

The brass carrier ring with bonded rubber for the outflow meter complies with the requirements of EN 13036-3:2003.

The brass ring is labelled with the manufacturer mark „BAM“ together with identification number and expiry date („verwendbar bis Monat/Jahr“).

Hardness according to ISO 868 of the ring number xx

(rubber ring measured on carrier): xx Shore A
Standard uncertainty: 1,0 Shore A

The listed values for measurement uncertainties have a confidence level of 95 % (coverage factor=1,96). Measurement uncertainties are determined from the combination of measurement error and the standard deviation of the measurement results, resulting from inherent batch fluctuations.

by order

Bundesanstalt für Materialforschung und -prüfung (BAM)

Dietmar Schulze; Quality Manager
Division 7.5
Unter den Eichen 87
12205 Berlin
Federal Republic of Germany

Tel.: +49 30 8104-3340
Fax: +49 30 8104-1707
Email: crm-elastomer@bam.de
Webshop: <https://www.webshop.bam.de/>