

date

The Reference Material (BAM-E007) delivered by BAM, Division 7.5 “Technical Properties of Polymeric Materials”, is prepared for test purposes and tested in accordance with DIN EN ISO/IEC 17025:2005.

The **sliders for the pendulum tester** comply with the requirements of EN 1097-8; chapter 7.3 und annex D.

Specification of the slider number XX as determined for the complete slider at 20 °C:

|   |                     |
|---|---------------------|
| Rebound resilience according to ISO 4662:<br>measurement uncertainty: | xx %<br>0,7 %       |
| Hardness according to ISO 48-2:<br>measurement uncertainty:           | xx IRHD<br>0,8 IRHD |

The mass of the slider complies with the requirement of  $(20 \pm 5)$  g.

The listed values for measurement uncertainties have a confidence level of 95 % (coverage factor=1,96).

The uncertainty of the values is furthermore influenced by material-related deviations within a production batch. Therefore in individual cases a total uncertainty of about a factor of two higher may result.

The aluminium backing is labelled with „BAM“ and together with identification number and date of manufacture (month, year).

by order

---

### **Bundesanstalt für Materialforschung und -prüfung (BAM)**

Dietmar Schulze; Quality Manager  
Division 7.5  
Unter den Eichen 87  
12205 Berlin  
Federal Republic of Germany

Tel.: +49 30 8104-3340  
Fax: +49 30 8104-1707  
Email: [crm-elastomer@bam.de](mailto:crm-elastomer@bam.de)  
Webshop: <https://www.webshop.bam.de/>