

date

The Reference Material (BAM-E017) delivered by BAM, Division 7.5 "Technical Properties of Polymeric Materials", is prepared for test purposes and tested in accordance with DIN EN ISO/IEC 17025:2005.

The properties of the rubber sheet **ISO 13226 SRE-NBR/L** correspond with the requirements given in ISO 13226.

The manufactured rubber reference plate is labelled (identification) as follows:

BAM	(producer)
ISO 13226	(standard)
SRE - NBR/L	(type of elastomer)
BAM-E017	(product number)
XXXX	(identification number)
19	(year of production)

Measurement results: determined from samples taken from a single standard plate from the charge mixture employed to produce plate number 155

	Requirements	Results	Standard Uncertainties
Increase in volume in liquid B for fuels, as in ISO 1817; 40 °C, 48 h	(53 - 59) %	54 %	± 0,9 %

The listed values for measurement uncertainties have a confidence level of 95 % (coverage factor=1,96). Measurement uncertainties are determined from the combination of measurement error and the standard deviation of the measurement results, resulting from inherent batch fluctuations.

by order

---

#### **Bundesanstalt für Materialforschung und -prüfung (BAM)**

Dietmar Schulze; Quality Manager  
Division 7.5  
Unter den Eichen 87  
12205 Berlin  
Federal Republic of Germany

Tel.: +49 30 8104-3340  
Fax: +49 30 8104-1707  
Email: [crm-elastomer@bam.de](mailto:crm-elastomer@bam.de)  
Webshop: <https://www.webshop.bam.de/>