

# CERTIFICATE OF ANALYSIS

## ERM®-BD513

Cd in Cocoa		
	Certified value <sup>1)</sup>	Uncertainty <sup>2)</sup>
Element	Mass fraction in mg/kg	
Cd	0.181	0.009

<sup>1)</sup> Unweighted mean value of the means of accepted sets of data, each set being obtained in a different laboratory and/or with a different method of determination. The certified value is traceable to the SI (Système International d'Unités) by the use of pure substances of known stoichiometry for calibration.

<sup>2)</sup> Estimated expanded uncertainty  $U$  with a coverage factor of  $k = 2$ , corresponding to a level of confidence of about 95%, as defined in the ISO/IEC Guide 98-3:2008 [Uncertainty of measurement -- Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)].

	Indicative Value <sup>1)</sup>	Uncertainty <sup>2)</sup>
Compound	Mass fraction in mg/kg	
Acrylamide	0.051	0.018

<sup>1)</sup> Unweighted mean value of the means of accepted sets of data, each set being obtained in a different laboratory and/or with a different method of determination. The informative value is traceable to the SI (Système International d'Unités) by the use of pure substances of known stoichiometry for calibration.

<sup>2)</sup> Estimated expanded uncertainty  $U$  with a coverage factor of  $k = 2$ , corresponding to a level of confidence of about 95%, as defined in the ISO/IEC Guide 98-3:2008 [Uncertainty of measurement -- Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)].

The property values will be valid for 12 months beginning with the dispatch of the material from BAM; this validity may be extended as further evidence of stability becomes available.

### NOTE

European Reference Material ERM®-BD513 was produced and certified under the responsibility of Bundesanstalt für Materialforschung und -prüfung (BAM) according to the principles laid down in the technical guidelines of the European Reference Materials® co-operation agreement between BAM-LGC-JRC.

Accepted as an ERM®, Berlin, 2019

Dr. S. Richter  
Committee for Certification

Dr. S. Recknagel  
Project Coordinator

Date of dispatch:

## ADDITIONAL MATERIAL INFORMATION

Determination of moisture using Karl-Fischer-titration performed at BAM gave the following non-certified results (mean of 45 single measurements  $\pm$  standard deviation):

**5.04 %  $\pm$  0.08 %**

The water content was seen to remain stable if the material is handled according to the instructions for use. The property values are valid for the material taken directly from the bag without drying prior to the analysis.

## DESCRIPTION OF THE SAMPLE

The material ERM<sup>®</sup>-BD513 is a powdered cocoa sample from commercial sources intended for human consumption. After mixing in a drum hoop mixer for 24 h the cocoa powder was filled into sealed aluminised plastic bags using a sample divider. Each plastic bag contains ca. 8 g of certified reference material.

## INTENDED USE

The intended purpose of reference material ERM<sup>®</sup>-BD513 is i) validation of analytical procedures for the determination of Cd and acrylamide in food, and ii) quality assurance in the analytical laboratory. The minimum sample size for chemical analysis is 0.5 g for Cd and 0.05 g for acrylamide.

## INSTRUCTIONS FOR USE

Before taking a sub-sample the material should be allowed to reach room temperature and is to be mixed thoroughly. Thereafter, the bag is to be closed tightly and stored at a temperature equal to or lower than -20 °C. To the best of our knowledge, the stability of the reference material is not affected by short periods of handling at ambient temperature during transport and use. However, BAM cannot be held responsible for any alterations of the material occurring during transportation to, and handling and storage at, the customer's premises, especially of opened samples.

## STORAGE

The material has to be stored at a temperature equal to or lower than -20 °C in its original plastic bag.

## PARTICIPANTS

Bundesamt für Verbraucherschutz und Lebensmittelsicherheit, Berlin, Germany  
Bundesanstalt für Materialforschung und -prüfung (BAM), Berlin, Germany  
Cork Public Analyst's Laboratory St. Finbarr's Hospital, Cork, Ireland  
CVUA Rheinland, Hürth, Germany  
Estonian Agricultural Research Centre, Saku, Estonia  
EXHM-GCSL/EIM Chemical Metrology Laboratory, Atnens, Greece  
Finnish Food Safety Authority, Helsinki, Finland  
Istituto Zooprofilattico Sperimentale del Piemonte, Liguria e Valle d'Aosta, Torino, Italia  
Kreis Mettmann Amt für Verbraucherschutz, Mettmann, Germany  
Laboratoire national de santé, Dudelange, France  
Landeslabor Schleswig-Holstein, Neumünster, Germany  
Landesuntersuchungsamt für Chemie, Hygiene und Veterinärmedizin, Bremen, Germany  
LUFÄ-ITL GmbH, Kiel, Germany  
National Food and Veterinary Risk Assessment Institute, Vilnius, Lithuania  
National Institute of Public Health - National Institute of Hygiene, Warsaw, Poland  
National Laboratory for Health, Environment and Food, Maribor, Slovenia  
Sanitary Veterinary and Food Safety Directorate, Bucharest, Rumania

## MEANS OF ACCEPTED DATA SETS

Means of accepted data sets,  
mass fraction in mg/kg

Line no.	Cd
1	0.150
2	0.151
3	0.159
4	0.160
5	0.164
6	0.168
7	0.174
8	0.175
9	0.182
10	0.187
11	0.187
12	0.188
13	0.194
14	0.196
15	0.198
16	0.200
17	0.200
18	0.204
19	0.210
<i>M</i>	0.181
<i>s<sub>M</sub></i>	0.019
$\sqrt{s_t^2}$	0.006

Means of accepted data sets,  
mass fraction in mg/kg

Line no.	Acrylamide
1	0.025
2	0.031
3	0.031
4	0.034
5	0.034
6	0.043
7	0.047
8	0.063
9	0.100
10	0.101
11	---
<i>M</i>	0.051
<i>s<sub>M</sub></i>	0.029
$\sqrt{s_t^2}$	0.008

The laboratory results have been examined statistically. Where a " --- " appears in the table it indicates that outlying results have been omitted.

*M* : mean of laboratory means

*s<sub>M</sub>* : standard deviation of laboratory means

$\sqrt{s_t^2}$  : averaged repeatability standard deviation (square root of the mean of laboratory variances)

### ANALYTICAL METHOD USED FOR CERTIFICATION

Element	Line no.	Method
Cd	1, 3, 9	ETAAS, dissolution in HNO <sub>3</sub> /H <sub>2</sub> O <sub>2</sub> , microwave
	2, 5, 13, 16, 17, 18, 19	ICP-MS, dissolution in HNO <sub>3</sub> /H <sub>2</sub> O <sub>2</sub> , microwave
	4, 14	ICP-MS, dissolution in HCl/HNO <sub>3</sub> , microwave
	6, 10, 11	ICP-MS, dissolution in HNO <sub>3</sub>
	7	ETAAS, dissolution in HNO <sub>3</sub> /HCl/H <sub>2</sub> O <sub>2</sub> , microwave
	8	HR-ICP-MS, dissolution in HNO <sub>3</sub> /H <sub>2</sub> O <sub>2</sub> , microwave
	12	ETAAS, dissolution in HNO <sub>3</sub> , microwave
	15	ID-ICP-MS, dissolution in HNO <sub>3</sub> /H <sub>2</sub> O <sub>2</sub> , HPA

#### Abbreviations:

ETAAS: Electrothermal atomic absorption spectrometry

ICP-MS: Inductively coupled plasma mass spectrometry

HPA: High pressure asher

HR-ICP-MS: High-resolution inductively coupled plasma mass spectrometry

Compound	Dataset no.	Method
Acrylamide	1, 3, 4, 5	LC-MS/MS acc. to EN 16618, D <sub>3</sub> -Acrylamide as ISTD
	2, 10	LC-MS/MS, D <sub>5</sub> -Acrylamide as ISTD
	6	GC-MS acc. to DIN 10785, D <sub>3</sub> -Acrylamide as ISTD
	7, 9	LC-MS/MS, D <sub>3</sub> -Acrylamide as ISTD
	8	LC-MS/MS, C <sub>13</sub> -Acrylamide as ISTD

**Abbreviations:**

GC:	Gas chromatography
ISTD:	Internal standard
LC:	Liquid chromatography
MS:	Mass spectrometry

**LEGAL NOTICE**

Neither BAM, its contractors nor any person acting on their behalf:

- (a) make any warranty or representation, express or implied, that the use of any information, material, apparatus, method or process disclosed in this document does not infringe any privately owned intellectual property rights; or
- (b) assume any liability with respect to, or for damages resulting from, the use of any information, material, apparatus, method or process disclosed in this document save for loss or damage arising solely and directly from the negligence of BAM.

**TECHNICAL REPORT**

A detailed technical report describing the production, general characterisation as well as the analysis procedures applied, and the treatment of the analytical data used to certify CRMs ERM<sup>®</sup>-BD513, ERM<sup>®</sup>-BD514 and ERM<sup>®</sup>-BD515 is available on request or can be downloaded from BAM website ([www.bam.de](http://www.bam.de)).

Supply of this Reference Material by:  
Bundesanstalt für Materialforschung und -prüfung (BAM)  
Richard-Willstätter-Str. 11, D-12489 Berlin, Germany

Phone: +49 30 8104 2061  
Fax: +49 30 8104 72061

E-mail: [sales.crm@bam.de](mailto:sales.crm@bam.de)  
Internet: [www.bam.de](http://www.bam.de)