

Certified Reference Material (CRM)

BAM-F023

Fluorescence Quantum Yield Standard

| Measurand ^{a)} | Certified value $\Phi_f^{b)}$ | Uncertainty $U^{c)}$ |
|---|-------------------------------|----------------------|
| BAM-F023 | 0.212 | 0.013 |
| <p>a) Ethanolic dye solution of BAM-F023, which shows an absorbance at the excitation wavelength $A(\lambda_{ex})$ between 0.11 and 0.135.</p> <p>b) BAM-F023 dissolved in absolute ethanol is certified for use as fluorescence quantum yield standard for the relative determination of the fluorescence quantum yield (Φ_f) and for controlling and validating the performance of integrating sphere-based fluorescence instruments for the absolute determination of Φ_f at $T = (22 \pm 2) \text{ }^\circ\text{C}$.</p> <p>c) Estimated expanded uncertainty U with a coverage factor of $k = 2$ corresponding to a level of confidence of approx. 95 %, as defined in the Guide to the expression of uncertainty.¹</p> | | |




Fig. 1 Normalized absorbance and fluorescence spectra of the ethanolic solution of BAM-F023 used for certification.

End of Validity

This certificate is valid for 2 years after dispatch, if the solid CRM is stored at 2 - 8 °C in the dark. BAM-F023 can be shipped at room temperature.

Sample No.:

Date of dispatch:

Material Description

The fluorescence quantum yield standard BAM-F023 is supplied as a solid in a tightly closed glass bottle^{x)}. The Standard Operation Procedure (SOP) based on the publication of C. Wuerth et al. in Nature Protocols 2013² that describes the correct use of CRM BAM-F023 is included. An open access version of the revised and finally submitted manuscript and supporting information approved by Nature Protocols can be found under <https://nbn-resolving.org/urn:nbn:de:kobv:b43-294202>.

^{x)}The bottle containing BAM-F023 and the box used as package have labels with different BAM logos as BAM meanwhile changed its logo.

Recommended Use

BAM-F023 dissolved in absolute ethanol is certified for use as a Φ_f standard for the relative determination of the fluorescence quantum yields of transparent luminophore solutions using excitation wavelengths between 480 nm and 635 nm (excitation within the dye's longest wavelength absorption band). BAM-F023 can be also employed for controlling and validating the performance of integrating sphere-based fluorescence instruments used for the absolute determination of Φ_f .

Handling

Please follow the instructions given in the Standard Operation Procedure (SOP) provided with the material.

Prerequisites for the reliable relative determination of Φ_f with BAM-F023 as reference standard is the use of calibrated spectrometers to determine the number of absorbed and emitted photons.³⁻⁴

Recommended instrument calibrations and characterizations include:

- a.) Calibration of the wavelength and intensity scale of the absorption spectrometer,
- b.) Calibration of the wavelength scale of the spectrofluorometer,
- c.) Determination of the range of linearity of the detectors of the spectrometers,
- d.) Determination of the wavelength-dependent relative spectral responsivity of the spectrofluorometer's detection channel and consideration of the photonic nature of the emitted light.

For the preparation of a dye stock solution, the solid dye should be dissolved in 10 mL absolute ethanol (spectrophotometric grade ≥ 99.8 % (GC)). Use of a freshly opened bottle of ethanol is recommended to guarantee a minimum water content.

For the relative determination of Φ_f in a 0° / 90° measurement geometry of the spectrofluorometer, an absorbance between 0.02 – 0.05 at the selected excitation

wavelength is recommended to minimize reabsorption effects.² For the absolute determination of Φ_f with integrating sphere-based fluorescence instruments, the choice of the optimum absorbance where reabsorption effects are minimum, depends on the size of the integrating sphere and the wavelength-dependent reflectivity of the sphere coating.³ The recommended measurement conditions are summarized in **Table 1**.

Table 1 Recommended measurement conditions.

| | |
|--|--|
| Measurement condition | |
| Excitation wavelength (λ_{ex}) | 480 nm - 635 nm |
| Absorbance at λ_{ex} | 0.02 - 0.05 (relative measurement) 0.03 - 0.15 (absolute measurement) |
| Temperature | 20 - 25 °C |

For further information, see the SOP provided by BAM and the comprehensive description of the procedures for the absolute and relative determination of transparent luminophore samples given in Nature Protocols.² An open access version of the revised and finally submitted manuscript and supporting information approved by Nature Protocols can be found under <https://nbn-resolving.org/urn:nbn:de:kobv:b43-294202>.

Transport and Storage

BAM-F023 should be shipped at room temperature. After receipt, the CRM must be stored in a refrigerator at 2 - 8 °C. Under these conditions, it will preserve its certified properties for **24 months after delivery**.

When preparing the measurement solution, care must be taken to avoid moisture uptake.

The ethanolic solution of BAM-F023 must be stored in the dark at 2 - 8 °C in a tightly closed container to prevent water uptake. Stock solutions of this dye with typical absorbances at the absorbance maximum of 2 can be stored under these conditions for six months in the dark at 2 - 8 °C. Dilute solutions with an absorbance at the maximum of 0.3 or less should be disposed after use.

BAM is not responsible for changes in the certified properties that occur during storage of the material at the customer's premises, especially not in the case of opened vials and prepared dye solutions.

Analytical Method used for Certification

The fluorescence quantum yield Φ_f is a direct measure for the efficiency of the conversion of absorbed light into emitted light. Φ_f is defined as the quotient of the number of emitted photons ($N_{em}(\lambda)$) and the number of absorbed photons ($N_{abs}(\lambda)$).

The absolute determination of the fluorescence quantum yield was performed with a calibrated custom-built integrating sphere setup at $T = (22 \pm 2) \text{ }^\circ\text{C}$. The measurement conditions used are given in Table 2.

Table 2 Measurement conditions used for certification: Excitation wavelength λ_{ex} , absorbances at the excitation wavelength $A(\lambda_{ex})$, wavelength of the maximum absorbance λ_{max} , maximum absorbance $A^s(\lambda_{max})$, and concentration c^s of a selected dye solution used.

| CRM | λ_{ex}/ nm | $A(\lambda_{ex})$ | λ_{max}/ nm | $A^s(\lambda_{max})$ | $c^s/ (\text{mol/L})$ |
|----------|---------------------------|-------------------|----------------------------|----------------------|-----------------------|
| BAM-F023 | 605 | 0.11 - 0.135 | 607 | 0.116 | 1.93E-06 |

With such a setup, the fluorescence quantum yield Φ_f can be directly measured as the quotient of the relative photon flux emitted from the sample (F_{em}) and the relative photon flux absorbed from the incident light at the chosen excitation wavelength (F_{abs}), see equation [1].

$$\Phi_f = \frac{F_{em}}{F_{abs}} \quad [1]$$

The relative absorbed photon flux F_{abs} and the relative emitted photon flux F_{em} are calculated by separating the measured spectra of sample and blank in an excitation and an emission region as shown in Figure 2. Subsequently, i) F_{abs} is calculated from the integrated difference of the spectrally corrected signals (I_c) of the blank (the solvent used for dissolving the dye) and the sample in the wavelength region of excitation (see equation [2]); ii.) F_{em} is determined from the integrated difference of the spectrally corrected signals of the sample and the blank in the spectral region of the emission (see equation [3]). Measurements with a blank are indicated with the index b, measurements with the sample with the index x, respectively. The emission correction curve (relative spectral responsivity $s(\lambda_{em})$, see equation [4]) was determined in reference to the spectral radiance and converted to the relative spectral photon flux (q) (see equation [2] and equation [3]).




Fig. 2 Scheme of the determination of the relative photon flux emitted from the sample (red area) and the relative absorbed photon flux (green area) measured with an integrating sphere setup.

$$F_{\text{abs}} = \int_{\lambda_{\text{ex1}}}^{\lambda_{\text{ex2}}} (q_{\text{b}}(\lambda_{\text{ex}}) - q_{\text{x}}(\lambda_{\text{ex}})) d\lambda_{\text{ex}} \quad [2]$$

$$= \int_{\lambda_{\text{ex1}}}^{\lambda_{\text{ex2}}} (I_{\text{c,b}}(\lambda_{\text{ex}}) - I_{\text{c,x}}(\lambda_{\text{ex}})) \lambda_{\text{ex}} d\lambda_{\text{ex}}$$

$$F_{\text{em}} = \int_{\lambda_{\text{em1}}}^{\lambda_{\text{em2}}} (q_{\text{x}}(\lambda_{\text{em}}) - q_{\text{b}}(\lambda_{\text{em}})) d\lambda_{\text{em}} \quad [3]$$

$$= \int_{\lambda_{\text{em1}}}^{\lambda_{\text{em2}}} (I_{\text{c,x}}(\lambda_{\text{em}}) - I_{\text{c,b}}(\lambda_{\text{em}})) \lambda_{\text{em}} d\lambda_{\text{em}}$$

For the certification measurements, 6 solid samples were randomly selected and dissolved in absolute ethanol. The absorbance of the measured dye solutions at the excitation wavelength of 605 nm was between 0.11 and 0.135. Each dye solution was measured 15 times.

Metrological Traceability

The certified fluorescence quantum yield Φ_{f} value of CRM BAM-F023 was determined with a calibrated custom-made integrating sphere setup. The calibration of the wavelength-dependent spectral responsivity of the integrating sphere setup was performed with the certified spectral radiance standard 10W BN-9701 (SDS, serial number 973372-02) calibrated by PTB Berlin in the wavelength region from 280 nm to 1750 nm (Calibration Certificate No. 73114 PTB 17; Oct. 2017), and with the deuterium lamp LSB 215 (D2, serial number H61685) from LOT Quantum Design GmbH, calibrated by Heraeus Noblelight GmbH for the wavelength range from 200 nm to 400 nm (Test report No.

21966_2). The spectral responsivity $s(\lambda)$ of the integrating sphere-detector ensemble was obtained by dividing the measured signals $\bar{I}^{raw}(\lambda)$ by the certified spectral radiance ($\text{Wm}^{-2}\text{sr}^{-1}$) of the SDS or deuterium lamp $Z_{RS}(\lambda)$ (see equation [4]), which are traceable to the International System of Units (SI).

$$s(\lambda) = \frac{\bar{I}^{raw}(\lambda)}{Z_{RS}(\lambda)} \quad [4]$$

References

1. GUM, Evaluation of measurement data – Guide to the expression of uncertainty in measurement. *Joint Committee for Guides in Metrology, Working Group 1* **2008**.
2. Wuerth, C.; Grabolle, M.; Pauli, J.; Spieles, M.; Resch-Genger, U., Relative and absolute determination of fluorescence quantum yields of transparent samples. *Nature Protocols* **2013**, *8* (8), 1535-1550.
3. Wuerth, C.; Lochmann, C.; Spieles, M.; Pauli, J.; Hoffmann, K.; Schuettrigkeit, T.; Franzl, T.; Resch-Genger, U., Evaluation of a Commercial Integrating Sphere Setup for the Determination of Absolute Photoluminescence Quantum Yields of Dilute Dye Solutions. *Appl. Spectrosc.* **2010**, *64* (7), 733-741.
4. Wuerth, C.; Pauli, J.; Lochmann, C.; Spieles, M.; Resch-Genger, U., Integrating Sphere Setup for the Traceable Measurement of Absolute Photoluminescence Quantum Yields in the Near Infrared. *Anal. Chem.* **2012**, *84* (3), 1345-1352.

Technical Report

A detailed technical report describing the analysis procedures and the treatment of the analytical data used to certify BAM-F023 is available upon request.

Legal Notice

Neither BAM, its contractors, nor any person acting on their behalf:

- (a) Make any warranty or representation, express or implied, that the use of any information, material, apparatus, method or process disclosed in this document does not infringe any privately-owned intellectual property rights; or
- (b) Assume any liability with respect to, or for damages resulting from the use of any information, material, apparatus, method, or process disclosed in this document save for loss or damage arising solely and directly from the negligence of BAM.

Accepted as BAM-CRM on July 1st, 2020

Bundesanstalt für Materialforschung und -prüfung (BAM)

Dr. S. Richter
Committee for Certification

Dr. J. Pauli/ Dr. U. Resch-Genger
Project Coordinators

Specimen

This reference material is supplied by:

Bundesanstalt für Materialforschung und -prüfung (BAM)
Richard-Willstätter-Str. 11, D-12489 Berlin, Germany

Phone : +49 30 8104 2061

Fax : +49 30 8104 72061

E-Mail: sales.crm@bam.de

Internet: www.webshop.bam.de